

## Certificate of Analysis

### Description

Product Name	NextSeq™ 550Dx High Output Reagent Kit v2.5 (75 cycles)	Catalog Number	20028870
Part Number	20028870	Lot Number	A148071

### Kit lot Contents

Material	Box Part Number	Box Lot number	Cartridge/ Tube Part Number	Cartridge/ Tube Lot Number
NextSeq 550Dx High Output Reagent Cartridge v2 (75 cycles)	20019560	A148071-1	20005419	20401275
NextSeq 550Dx Buffer Cartridge v2 (75 cycles)	20019561	A148071-2	20005420	20396367
NextSeq 550Dx High Output Flowcell Cartridge v2.5 (75cycles)	20031098	A148071-3	20026363	20394570
NextSeq 550Dx Accessory Box (75 cycles)	20019563	A148071-4	20018864	20396783

### Test Conditions

Kitted reagents were tested on a NextSeq 550Dx sequencing system in a 2x38 cycle paired end run configuration with PhiX at a concentration which produced a cluster density of 160-230 K/mm<sup>2</sup>. Flow Cells included in the kit lot were manufactured and released in accordance with production specifications.

### Test Results

Metric	Specification	UOM	Result
Sequencing output	≥ 22.5	Gigabases	Pass
Cumulative Q-score <sup>1</sup> ≥ 30	≥ 80	%	Pass

<sup>1</sup> Q-Scores measure the probability that a base is called incorrectly. A higher quality score indicates a smaller probability of error. A quality score of 30 represents an error rate of 1 in 1000, with a corresponding call accuracy of 99.9%.

### Certification

This document certifies that the product(s) described above meet quality specifications.

#### Quality Review

Print Name	Tang Wei Qiang	Signature		Date	25-OCT-2019
------------	----------------	-----------	---	------	-------------